

C 150 - The new STANDARD in CO2-incubators

The C 150 sets new standards for cell cultivation: contamination-free as a result of hot air sterilization at 180 °C, stable pH thanks to a drift-free FPI infrared measurement system, superior temperature precision with excellent dynamics, and an interior that is condensation-free even at high humidity levels to ensure best growth results. Used stand alone or as a stack the C 150 is easily accessible.



▶ Performance features and equipment :

- Temperature range 7 °C (13 °F) above ambient up to 50 °C (122 °F)
- Air jacket system assuring temperature accuracy and reproducible results
- Water pan with built-in condensation control maintains dry interior walls up to 95 % RH
- Standard compliant 180 °C (356 °F) hot air sterilization using an overnight cycletime (acc to. DIN 58947, Pharmacopoeias, ANSI)
- Weldless deep-drawn inner chamber made of stainless steel with integrated shelf support system
- Drift free infrared CO2 measurement system
- Microprocessor with LED display for temperature and CO2-concentration
- Various alarm and status displays
- Automatic diagnostic system with optical and audible alarm, as well as zero-voltage relay contact for central monitoring
- Lockable controller keyboard via 3 digit password
- Independent adjustable temperature safety device class 3.1, providing full protection against chamber over-temperature, with visual and audible temperature alarm
- Gas mixing head
- Tightly-fitting inner glass door
- Lockable door
- Door hinged left or right, factory fitted
- Units are stackable with stacking adapter
- 3 perforated shelves made of stainless steel
- BINDER test certificate
- Two BINDER CO2-incubators, stacked:
- Two incubators C 150 series, incl. stacking adapter, ensuring direct thermal decoupled stacking



C 150

Exterior dimensions	
Width (mm/inch)	680 / 26.8
Height (mm/inch)	819 / 32.2
Depth (mm/inch)	815 / 32.1
plus door handle, l-triangle (mm/inch)	54 / 2.1
plus mains- and gas connection (mm/inch)	60 / 2.4
Wall clearance rear (mm/inch)	100 / 3.9
Wall clearance side (mm/inch)	50 / 2.0
Interior dimensions	
Width (mm/inch)	500 / 19.7
Height (mm/inch)	600 / 23.6
Depth (mm/inch)	500 / 19.7
Interior volume (l/cu.ft.)	150 / 5.4
Perforated shelves, stainless steel (number standard/max.)	3 / 6
Dimensions of perforated shelves, width x depth (mm)	491 x 442
width x depth (inch)	19.3 x 17.4
Weight of the unit (kg/lbs.)	95 / 209
Temperature / CO2 data	
Temperature range, 7 °C (44,6 °F) above ambient up to °C / °F	50
Temperature variation at 37 °C (98.6 °F) (± °C)	0.4
Temperature fluctuation (± °C)	0.1
Recovery time after door was opened for 30 sec. at 37°C (98.6°F) (Min) 1)	5
CO2 data	
CO2-range (Vol.-% CO2)	0 - 20
Setting accuracy (Vol.-% CO2)	0.1
Recovery time after door was opened for 30 sec. up to 5 vol. % (Min) 1) 2)	7
CO2 measurement	IR
Hose connectors for CO2 (mm/inch)	6
Humidity data	
Humidity (average values) (±2 % RH)	95 (±2)
Electrical data	
Housing protection acc. to EN 60529	IP 20
Nominal voltage (±10 %) 50 / 60 Hz (V)	230
Nominal power (kW)	1.4
Energy consumption 2) at 37°C (W)	110

1) up to 98 % of the set value

2) The recovery times of the gas concentration inside the chamber following a door opening refer to an inlet pressure of 2.0 bars (psi).

All technical data is specified for units with standard equipment at an ambient temperature of 25 °C (77 °F) and a mains voltage fluctuation of +5%/-10%. The temperature data is determined in accordance to factory standard following DIN 12880, observing the recommended wall clearances of 10 % of the height, width, and depth of the inner chamber. Technical data refers to 100 % fan speed. All indications are average values, typical for units produced in series. We reserve the right to alter technical specifications at all times.



► Cleaning kit

Halogenfree detergents for gentle cleaning of interior chamber without causing corrosion:

- Mild cleaner for stainless steel to provide long -therm-protection
- Clinically approved disinfectant
- Cleaning cloth



► Calibration certificate for temperature and CO2

Factory calibration certificate for temperature and CO2. Temperature measurement in center / CO2 measurement performed using analyzed test gas at 37 °C (98.6 °F) and 5 % CO2.

**C 150**

Perforated shelf, stainless steel	<input type="radio"/>
Base on castors	<input type="radio"/>
Stacking adapter for direct thermal decoupled stacking of a CB / C 150 combination	<input type="radio"/>
Stacking adapter for direct thermal decoupled stacking of two C 150 models	<input type="radio"/>
Factory calibration certificate for temperature and CO ₂ . Temperature measurement in center / CO ₂ measurement performed using analyzed test gas at 37 °C (98.6 °F) and 5 % CO ₂ .	<input type="radio"/>
Gas Supply Control Unit, for the connection of 2 gastank, either CO ₂ or N ₂ , with acoustic and visible alarm. (Available 2. quarter 2009)	<input type="radio"/>
Temperature measurement acc. to DIN 12880 (27 measuring points) at 37 °C (98.6 °F) or at specified temperature with measuring protocol and certificate	<input type="radio"/>
Gastank connection kit for CO ₂ , consisting of a gastank pressure regulator, max. pressure 10 bar, with connection parts, and a 5 m (16.4 ft.) hose	<input type="radio"/>
Manual for Primary Human Cell Culture, in English	<input type="radio"/>
Cleaning kit for equipment maintenance and disinfection, consisting of pH - neutral detergent (1000 g / 2.21 lbs., concentrate), spray disinfectant (500 ml / 0.13 gal. ready - for - use solution) and lint - free cleansing tissues	<input type="radio"/>



Thank you for reading this data sheet.

For pricing or for further information, please contact us at our UK Office, using the details below.



UK Office

Keison Products,

P.O. Box 2124, Chelmsford, Essex, CM1 3UP, England.

Tel: +44 (0)1245 600560

Fax: +44 (0)1245 600030

Email: sales@keison.co.uk

Please note - Product designs and specifications are subject to change without notice. The user is responsible for determining the suitability of this product.