

## Bridgman Crystal Growth Furnace - BV-HTRV

### General Information

The BV-HTRV is as a special crystal growing furnace according to the Bridgman method.

The Bridgman method uses a pre-synthesized material that moves slowly through a temperature gradient. The melted material moves through a decreasing temperature gradient and forms a single crystal. The BV-HTRV is a tube furnace which is mounted on a device engineered specifically for the Bridgman method. As a standard, the HTRV 70-250 or the HTRV 100-250 are used as a tube furnace mounted on the pulling device. In principal, every tube furnace can be mounted on the pulling device. Most common are the two models, HTRV 70-250 and the HTRV 100-250.

The short heated length is an advantage as it creates an ideal gradient for the Bridgman method. The temperature decreases towards the bottom of the furnace. The pulling device moves the sample with an adjustable speed toward the lower temperature. Next to the sample is a thermocouple in order for an accurate reading of the sample temperature. Both, sample and probe thermocouple are attached to the bottom pulling device. The movement of the sample can be fast for loading and unloading or with a user defined speed for crystal growth.

A ceramic tube surrounds the probe thermocouple and sample. At both ends of the tube, a water cooled flange is connected. At the top of the system, the tube and flange are fixed. A bellow connects the tube and pulling device at the bottom of the furnace. The bellow is extended during the downward movement of the sample. Operation of the unit can be achieved under vacuum conditions.

At the top, the tube is connected to the vacuum pump. The valve towards the vacuum line is opened and closed manually. The vacuum level is controlled by a piezo measurement gauge. A manually operated rotameter allows purging of inert gas. To reduce the Oxygen value prior to the crystal growing process, evacuation and backflooding with an inert gas is performed several times.

It is possible to connect a computer to the system in order to log all the relevant data of the process, i. e. position of the sample and temperature of the probe thermocouple. To load and unload the sample, the clamps must be opened. With the fast moving option, the sample is easily accessible.

The pulling device that enables Bridgman type crystal growth can be combined with all single and multi-zone tube furnaces.

### Standard features

- Bridgman method - Crystal growing
- In vacuum up to 1450 °C
- In an inert atmosphere up to 1800 °C
- Precisely defined and controlled pulling speed
- Manual operation
- Data recording option



## **Bridgman Crystal Growth Furnace - BV-HTRV**

### **Technical Specifications**

#### **BV-HTRV 70-250**

Tmax (°C)	1600, 1700, 1800
Max. outer diameter accessory tube (mm)	100
Heated length (mm)	250
Furnace dimensions H x W x D (mm)	1800 x 950 x 750
Furnace weight (kg)	300
Control module dimensions H x W x D (mm)	850 x 560 x 600
Control module weight (kg)	60
Power (kW)	6.5



Thank you for reading this data sheet.

For pricing or for further information, please contact us at our UK Office, using the details below.



**UK Office**

**Keison Products,**

**P.O. Box 2124, Chelmsford, Essex, CM1 3UP, England.**

**Tel: +44 (0)330 088 0560**

**Fax: +44 (0)1245 808399**

**Email: [sales@keison.co.uk](mailto:sales@keison.co.uk)**

Please note - Product designs and specifications are subject to change without notice. The user is responsible for determining the suitability of this product.